

ASTM D 97 - D 2500 - ISO 3015 & ISO 3016 - IP 15 & IP 219

SCOPE

Standard Test Method for Pour Point of Petroleum Products

CLOUD AND POUR POINT BATH

- **CPP BASIC**-

- **REF 941584**

MAIN CHARACTERISTICS

- ✓ Temperature bath with 5 wells
- ✓ Stirrer
- ✓ One set of 6 thermometers 5C
- ✓ Cooled by dry ice or cryostat

SCOPE OF DELIVERY

5 felt discs sold in packs of 10 under **REF 9415901**

5 felt rings sold in packs of 10 under **REF 9415902**

REF 19439 Test tube
REF 11492 ASTM thermometer (5C)

NECESSARY ACCESSORIES

REF 941597 Dry Ice Machine
REF 22237 Ultra Cryostat (-81/+100°C ±0.02)
REF 21131 Insulated pipe for cryostat internal Ø 7mm
REF 22216 Cooling cryostat (-40/+200°C ±0.02)
REF 11493 ASTM thermometer (6C)
REF 11530 ASTM thermometer (61C)

REFERENCE FLUIDS

REF 60002 Cloud point (Diesel) ; -5,3°C
REF 60003 Pour Point (Diesel) ; -31,7°C



ORDERING INFORMATION

REF 941584

For use on AC 230 V, 50 Hz – 1 A

(W) 290x (D) 360 x (H) 250 mm (± 9 kg)

CONTACT : sales@normalab.com

NORMALAB FRANCE SAS
ZA Caux Multipôles 1 - F-76 190 Valliquerville
Tel. : +33 232 700 100
Fax : +33 232 704 732

www.normalab.com

CPP BASIC Leaflet 15

DISTRIBUTED BY